

HIGH-SENSITIVITY VIBRATIONAL SPECTROSCOPY BY CANTILEVER ENHANCED PHOTOACOUSTIC SENSOR

JACOPO PELINI, *University of Naples "Federico II", Corso Umberto I 40 - 80138 Naples, CNR-INO, and LENS, 50019 Sesto Fiorentino, Italy*; **MARIO SICILIANI de CUMIS**, *ASI, Località Terlecchia, 75100 Matera, Italy*; **ZHEN WANG**, *The Chinese University of Hong Kong, New Territories, Hong Kong*; **IACOPO GALLI, INAKI LOPEZ GARCIA, PABLO CANCIO PASTOR**, *CNR-INO, and LENS, 50019 Sesto Fiorentino, Italy*; **NAOTA AKIKUSA**, *Hamamatsu Photonics K.K., Shizuoka, Japan*; **MARIA CONCETTA CANINO**, *CNR-IMM, Via P. Gobetti, 101 - 40129 Bologna, Italy*; **WEI REN**, *The Chinese University of Hong Kong, New Territories, Hong Kong*; **PAOLO DE NATALE, SIMONE BORRI**, *CNR-INO, and LENS, 50019 Sesto Fiorentino, Italy*

The race towards compact and robust sensors able to detect extremely low concentrations of molecules in the air plays an important role in our modern society, impacting sectors such as environmental and climate change monitoring, transportation, agriculture, safety, and security. During the last decade, optical detection with ultra-high sensitivity, down to the part-per-quadrillion level, was demonstrated with cavity-ring down techniques¹. An alternative approach deals with photoacoustic spectroscopy, in which the target gas selectively absorbs a modulated laser light and generates acoustic waves by non-radiative energy relaxation processes, which are then converted into a measurable signal by a proper acoustic transducer. This approach is particularly appealing for its unique characteristics of robustness, record dynamic range, and compact size, which make them particularly attractive for in-field applications. Moreover, photoacoustic-based sensors have shown great potential in achieving a sensitivity at the level of the techniques mentioned above, when combined with narrow-linewidth mid-infrared lasers and high-finesse optical cavities^{2,3,4,5}.

¹M.G. Delli Santi et al., *Biogenic Fraction Determination in Fuel Blends by Laser-Based 14CO₂ Detection*, *Advanced Photonics Research* 2, 2000069 (2021)

²G. Zhao et al., *High-resolution trace gas detection by sub-Doppler noise-immune cavity-enhanced optical heterodyne molecular spectrometry*, *Opt. Express* 27, 17940 (2019)

³S. Borri et al., *Intracavity quartz-enhanced photoacoustic sensor*, *Appl. Phys. Lett.* 104, 091114

⁴Z. Wang et al., *Doubly resonant sub-ppt photoacoustic gas detection with eight decades dynamic range*, *Photoacoustics* 27, 100387 (2022)

⁵T. Tomberg et al., *Sub-parts-per-trillion level sensitivity in trace gas detection by cantilever-enhanced photo-acoustic spectroscopy*, *Sci. Rep.* 8,1848 (2018)

In this framework, we designed a novel sensor combining all the advantages of the photoacoustic technique with the sensitivity enhancement provided by a doubly-resonant acoustic-optic system. In our setup a 4.5 μ m-wavelength CW quantum cascade laser excites strong N₂O and CO rovibrational transitions. The acoustic-to-voltage transduction of the signal is performed by a Micro Electro Mechanical cantilever, whose displacement is measured with a balanced Michelson interferometer. The cantilever is mounted in a homemade photoacoustic cell consisting of a high-Q-factor acoustic resonator placed inside a high-finesse optical resonator. This design, leveraging on a double standing wave effect, achieves a combined acoustic and optical amplification factor of several orders of magnitude compared to the standard configuration, thus strongly enhancing the final detection sensitivity.